

GEOSTATS PTY LTD

Mining Industry Consultants
Reference Material Manufacture and Sales

Certified Geochem Base Metal Reference Material Product Code

GBM303-6

Certified Control Values



GBM303-6

Geostats Pty Ltd, Certified Geochem Base Metal Reference Material, Product Code:

Element	Grade	Standard Deviation	Num of Analyses	Confidence Interval
Nickel (ppm)	106	10	51	+/- 2.8
Copper (ppm)	13967	549	66	+/- 135.9
Zinc (ppm)	57	8	49	+/- 2.4
Lead (ppm)	56	11	52	+/- 3.1
Arsenic (ppm)	603	29	47	+/- 8.5
Cobalt (ppm)	151	16	54	+/- 4.5
Silver (ppm)	5.5	0.8	57	+/- 0.2

CRM Details

<u>Control Statistic Details</u> Control statistics were produced from results accumulated in the April-2003 round robin. The number of results used to certify each analyte is shown in the table above.	<u>Neutron Activation Analysis Results (ppm, unless otherwise noted)</u>		<u>Major Elements by Fusion / XRF (%)</u>	
		Antimony	0.81	Fe
	Arsenic	610	SiO ₂	nr
	Barium	227	Al ₂ O ₃	nr
<u>Material Description</u> This material is described as a Sulphide Cu / Au Ore Pilbarra.	Bromine	3.52	TiO ₂	nr
	Cadmium	nr	MnO	nr
	Caesium	<1	CaO	nr
	Calcium (%)	2.77	P	nr
	Cerium	134	S	nr
	Chromium	56.5	MgO	nr
	Cobalt	154	K ₂ O	nr
	Europium	0.82	Na ₂ O	nr
	Gold (ppb)	6640	LOI1000	nr
	Hafnium	3.22		
	Iridium (ppb)	<20	Neutron Activation Analyses and Fusion / XRF Analyses are single results and are indicative only. These are provided for matrix identification purposes.	
	Iron (%)	10.9	nr: Not Reported	
	Lanthanum	84.3		
	Lutetium	0.65		
	Mercury	nr		
	Molybdenum	36.8		
	Neodymium	nr		
	Nickel	nr		
	Potassium (%)	1.41		
	Rubidium	98.8		
	Samarium	9.74		
	Scandium	13.1		
	Selenium	<5		
	Silver	<5		
	Sodium (%)	0.677		
	Strontium	nr		
	Tantalum	<1		
	Tellurium	8.8		
	Terbium	nr		
	Thorium	12.7		
	Tin	nr		
	Tungsten	1140		
	Uranium	<2		
	Ytterbium	4.96		
	Zinc	<100		
	Zirconium	<500		
<u>Colour Designation (ISCC-NBS, SP440)</u> This material is dark gray in colour.				
<u>Usage</u> This product is for use in the mining industry as a reference material for monitoring and testing the accuracy of laboratory assaying.				
<u>Preparation and Packaging</u> All CRMs are dried in an oven for a minimum of 12 hours at 110°C. The dry material is then pulverised to better than 75 micron (nominal mean of 45 micron) using an air classifier. The material is then homogenised and stored in a sealed, stable container ready for final packaging. Materials are statistically sampled from stores, then packaged into either heat sealed, air tight, plastic pulp packets or screw top sealed plastic containers ready for distribution. All packaging has been chosen to ensure minimal contamination from outside sources during shipment, use and storage.				
<u>Assay Testwork</u> All standards are tested thoroughly in the Geostats bi-annual laboratory survey. This involves assaying by multiple laboratories from around the world. Results are compiled into a comprehensive report detailing statistics for each standard. Assay distributions are checked and processed statistically, producing monitoring statistics for these standards. Materials are tested regularly to ensure stability and homogeneity.				
<u>Stability</u> This product remains stable in its original packaging, away from direct sunlight.				
<u>Material Safety</u> This product is not hazardous and non-toxic.				

20 Hines Road, O'Connor, Western Australia 6163

Phone : +61 8 9314 2566, Fax : +61 8 9314 3699

e-mail : pjh@geostats.com.au, srr@geostats.com.au

Website <http://www.geostats.com.au>