

GEOSTATS PTY LTD

Mining Industry Consultants
Reference Material Manufacture and Sales

Certified Multi-Element Reference Material Product Code

GBMS623-3

Certified Control Values

Analyses

Element	Grade	Standard Deviation	No of Analyses	Confidence Interval
Au - FA (ppm)	5.85	0.39	99	+/- 0.079
Au - AR (ppm)	5.46	0.44	64	+/- 0.112
Silver (ppm)	3.7	0.8	181	+/- 0.12
Copper (ppm)	10792	537	204	+/- 74.3
Lead (ppm)	322	24	187	+/- 3.5
Zinc (ppm)	35	7	184	+/- 1.1
Nickel (ppm)	77	10	183	+/- 1.5
Arsenic (ppm)	570	54	177	+/- 8
Cobalt (ppm)	332	30	175	+/- 4.6
Sulphur (%)	5.16	0.22	131	+/- 0.038

CRM Details

Control Statistic Details

Control statistics were produced from results accumulated in the :

Oct-98, Apr-00, Apr-06 Geostats Pty Ltd Laboratory Round Robin Programs.
64 laboratories (at least) tested this material for base metal content.

Source Material

Prior to homogenisation and testing, this material was sourced from Copper/Gold ore oxide ex Pilbara region.

Colour Designation

Light brownish gray

Usage

This product is for use in the mining industry as reference materials for monitoring and testing the accuracy of laboratory assaying.

Preparation and Packaging

All standards are dried in an oven for a minimum of 12 hours at 110C. The dry material is then pulverised to better than 75 micron (nominal mean of 45 micron) using an Air Classifier. The material is then homogenised and stored in a sealed, stable container ready for final packaging.

Materials are statistically sampled from stores, then packaged into either heat sealed, air tight, plastic pulp packets or screw top sealed plastic containers ready for distribution. All packaging has been chosen to ensure minimal contamination from outside sources during shipment, use and storage.

Assay Testwork

All standards are tested thoroughly in the Geostats bi-annual laboratory survey. This involves assaying by a minimum of 50 reputable laboratories selected from across the world using a variety of methods (including FA, AR, 3AD, 4AD and ICP, AAS and XRF). Results are compiled into a comprehensive report detailing statistics for each standard. Assay distributions are checked and processed statistically, producing monitoring statistics for these standards. Materials are tested regularly to ensure stability and

Neutron Activation


Analysis Results (ppm)

Antimony	2
Arsenic	558
Barium	<100
Bromine	7
Cadmium	nr
Cerium	206
Caesium	<1
Chromium	496
Cobalt	356
Europium	1
Gold ppb	5400
Hafnium	7
Iridium ppb	<20
Iron %	6
Lanthanum	113
Lutetium	1
Molybdenum	<5
Nickel	nr
Rubidium	41
Samarium	14
Scandium	22
Selenium	<5
Sodium %	1
Tantalum	<1
Tellurium	<5
Terbium	nr
Thorium	10
Tin	nr
Tungsten	129
Uranium	<2
Ytterbium	4
Zinc	<100
Zirconium	<500
Calcium%	8
Potassium %	1
Silver	<5
Mercury	nr
Neodymium	nr
Strontium	nr

Major Elements

Fusion / XRF (%)

Fe	nr
SiO ₂	nr
Al ₂ O ₃	nr
TiO ₂	nr
MnO	nr
CaO	nr
P	nr
S	nr
MgO	nr
K ₂ O	nr
Na ₂ O	nr
LOI1000	nr



20 Hines Road, O'Connor, Western Australia 6163

Phone : +61 8 9314 2566

e-mail : info@geostats.com.au

Website <http://www.geostats.com.au>

GBMS623-3

Geostats Pty Ltd, Certified Multi-Element Reference Material, Product Code :