

GEOSTATS PTY LTD
Mining Industry Consultants
Reference Material Manufacture and Sales

Certified Antimony Reference Material Product Code

GSB-02

Certified Control Values

Element	Grade	Standard Deviation	No of Analyses	Confidence Interval
Antimony (ppm)	310838	10206	50	+/- 2930
Gold (ppm)	23.64	0.72	38	+/- 0.24

CRM Details

Control Statistic Details

Control values for this material were determined during a dedicated certification program in December-2014.

Analyses on this material are expected to be within 3 times the standard deviation of the average grade.

The confidence interval is an indication of the quality of testwork on the material and the quality of the material itself, not to be confused with the control limits for assaying.

Description of Source Material

This material is described as a Composite of plant concentrate and tails.

Usage

This product is for use in the mining industry as a reference material for monitoring and testing the accuracy of laboratory assaying.

Preparation and Packaging

All standards are dried in an oven for a minimum of 6 hours at 105C. The dry material is then pulverised to better than 95% passing 75 micron using a puck and bowl pulveriser. The material is then homogenised and stored in a sealed, stable container ready for final packaging.

Materials are statistically sampled from stores, then packaged into heat sealed, air tight, plastic packets ready for distribution. All packaging has been chosen to ensure minimal contamination from outside sources during shipment, use and storage.

Assay Testwork

This reference material was tested in a dedicated certification program. 10 pulp sub-samples were sent to 5 laboratories for analysis. Assay distributions were checked and processed statistically, producing monitoring statistics for this reference material. Samples of the material are tested regularly to ensure stability and homogeneity.

Stability

This product remains stable in its original packaging, away from direct sunlight.

Material Safety

This product is not hazardous and non-toxic.

20 Hines Road, O'Connor, Western Australia 6163
Phone : +61 8 9314 2566, Fax : +61 8 9314 3699
e-mail : pjh@geostats.com.au, srr@geostats.com.au
Website <http://www.geostats.com.au>

GSB-02

Geostats Pty Ltd, Certified Antimony Reference Material, Product Code :